APPARATUS AND METHODS FOR
DIAGNOSING OSTEOPOROSIS AND OTHER DISEASES
WITH MR IMAGING

5   **FIELD OF THE INVENTION**

This invention generally relates to magnetic resonance imaging (MRI) utilizing nuclear magnetic resonance (NMR) phenomena. It more particularly relates to a specialized MRI system used for diagnosis of osteoporosis and other diseases.

10

BACKGROUND OF THE INVENTION

Osteoporosis is a widespread disease characterized by bone loss usually associated with aging. It is a complex and chronic disease that may not be recognized until resulting fractures late in life. 15 In many countries around 50% of women above age of 70 is osteoporosis and to overcome it has become an important social requirement.

For diagnosis and follow-up of osteoporosis, measurement of bone mineral density (BMD) is indispensable. The conventional methods for the BMD measurement are based on X-ray attenuation measurements 20 through a bone. Among them, DEXA (Dual-Energy X-ray Absorptiometry) is now the standard method, but cannot be used repeatedly nor for young women because of the ionizing radiation. Thus the X-ray BMD measurement should be replaced by some noninvasive and high precision method such as bone marrow volume measurement using MRI as described 25 below, because MRI has no ionizing radiation and spatial localization capability is guaranteed.

The bones of a human skeletal system have a dense outer shell made

of cortical bone (alternatively termed "compact" or "dense" bone). Inside the cortical bone, many regions have a mesh or network of trabecular bone (alternatively termed "spongy" or "cancellous" bone), made up of roughly the same material as the cortical bone. The regions enclosed by the cortical bone and/or trabecular network are filled with "bone marrow", of which major components are water and lipid. Since the protons (hydrogen nuclei) of the water and lipid can be measured using MRI, the bone marrow volume can be quantified using MR images. Thus quantitative bone MRI can be a complementary method for the measurement of BMD because bone is almost chemically homogeneous. Studies according to this concept are found in the following references: Jara el al., "High-Resolution Variable Flip Angle 3D MR Imaging of Trabecular Microstructure in Vivo", Magnetic Resonance in Medicine **29**: 528-539, 1993; Wehrli et al., "Cancellous Bone Volume and Structure in the Forearm: Noninvasive Assessment with MR Microimaging and Image Processing", Radiology **206**: 347-357, 1998; L. Hilaire et al., "High-Speed Spectroscopic Imaging for Cancellous Bone Marrow R_2* mapping and Lipid Quantification", Magnetic Resonance Imaging, **18**: 777-786, 2000.

MRI is by now a commercially available and widely accepted non-invasive method for measuring information about internal structures of living systems and heterogeneous materials. However, most of medical MRI are developed for a whole human body and require a large cost and space. For the measurement of bone marrow volume, a compact MRI system for a part of a human body is appropriate because bone density for a whole human body can be well estimated from a measurement of a part of bone.

SUMMARY OF THE INVENTION

The most serious fracture for osteoporosis patients is the hip fracture because the patients cannot move by themselves for more than several months. To acquire MR images of a hip bone (femur) for diagnosis of osteoporosis, a whole body MRI system is indispensable, requiring a large cost for the patients. However, because correlations among densities of bones at various parts of a human body are very high, the bone density at the femur can be well estimated from those of other bones.

In this invention, calcaneus is used for the bone density measurement using MRI because of the following reasons: first, calcaneus is made of homogeneous trabecular bone and reflects the whole-body bone status very well; second, calcaneus is a peripheral part of a human body and easy to measure using a compact MRI system; third, the magnet, usually the heaviest unit of an MRI system, can be placed in the lower level and the MRI system becomes mechanically stable. For the measurement of a heel, an RF probe box having an oval aperture with a support pad to fix the heel in the RF coil is used. In the RF coil, external reference materials are placed and imaged together with the heel for NMR signal intensity calibration.

Quantification of materials using a compact MRI system gives rise to several new problems, because the magnet, gradient coil, and RF coil are small compared to those of a whole body MRI and the inhomogeneity of the static magnetic field, magnetic field gradients, and RF magnetic field become large compared to that of a whole body

MRI. To measure proton density of bone marrow of calcaneus in such inhomogeneous magnetic fields, spin-echo imaging sequences with different echo times are used to image a heel together with external reference materials. By combining measurements with different echo

5   times and reference materials, T_2 decay effects and spatially varying image intensities are corrected. As a result, the proton density of bone marrow is obtained and used for evaluation and follow-up of osteoporosis and other diseases.

10   **BRIEF DESCRIPTIONS OF THE DRAWINGS**

FIG.1 is a schematic overview of a compact MRI system for imaging of a heel;

FIG.2 is a schematic representation of an RF probe box with an oval aperture used for imaging of a heel;

15   **FIG.3** schematically depicts a solenoid coil and matching circuit for the RF probe;

FIG.4 is a schematic cross-sectional depiction of a support pad for a heel and external reference materials, inserted in the open-bore of the RF probe shown in FIG.2;

20   **FIG.5** is a schematic cross-sectional depiction of another type of the support pad for a heel and external reference material, inserted in the open-bore of the RF probe shown in FIG.2;

FIG.6 is a standard 2D spin-echo pulse sequence used to quantify the proton density of bone marrow in calcaneus;

25   **FIG.7** is a flow diagram of one method for calculating proton density using the MRI system shown in FIG.1;

FIG.8 schematically shows a vertical cross-sectional image of a heel

with external reference phantoms;

FIG.9 schematically shows a vertical cross-sectional image of a water phantom for intensity correction of images of heels with external reference phantoms;

5 **FIG.10** is a semilogarithmic plot of spin-echo image intensities of a plant oil phantom plotted against the echo time;

FIG.11 is a large flip-angle excitation 2D spin-echo pulse sequence to shorten the repetition time for the quantification of the proton density of bone marrow in calcaneus.

10

DETAILED DESCRIPTION OF THE INVENTION

FIG.1 schematically illustrates an MRI system with a permanent magnet, gradient coil, RF probe, and MRI console for imaging of a 15 heel, as an exemplary embodiment of this invention.

In **FIG.1**, the heel **1** is inserted into the RF coil **2** placed between the permanent magnet **3** and gradient coil **4**. The permanent magnet produces a homogeneous magnetic field of about 0.21T over the 12 cm diameter spherical volume in the central region of the magnet 20 gap, of which width is 16 cm. The volume of the homogeneous magnetic field is sufficient to image a heel and external reference phantoms at the same time. The RF coil is actually stored in an RF shielded box as shown in **FIG.2** to eliminate the electromagnetic coupling between the RF coil **2** and gradient coil **4**.

25 MRI measurements are controlled by the MRI console, which consists of a display monitor **5**, computer system **6**, RF transceiver **7**, gradient driver **8**, and RF transmitter **9**. The computer system includes a digital

signal processor (DSP) board for the MRI pulse programmer, a direct digital synthesizer (DDS) board for generation of a radio frequency source for the NMR transceiver, and an analog-to-digital converter (ADC) board for digitization of the NMR signal.

5   The RF transceiver **7** is used to generate RF pulses for excitation of the proton spins of the heel and convert the NMR signal modulated at the Larmor frequency at around 8.9 MHz to a lower frequency (several tens of kHz) having spatial information of the proton spins. The gradient driver **8** is used to supply electric currents to the gradient

10   coil **4** according to waveforms generated by the pulse programmer. The transmitter **9** is used to supply RF currents to the RF coil **2**.

FIG.2 schematically illustrate the RF probe used in the system. The RF probe box **10** is made of brass plates with an oval aperture **11** in the upper side. The dimension of the box is 22 cm (width) x 16

15   cm (height) x 13 cm (depth) and that of the aperture is 16 cm (long axis) x 8 cm (short axis). The aperture can accommodate heels of most female adult patients as depicted as **12**. The probe has an open-bore RF coil support **13** made of acrylic plates, which goes from the upper aperture down to another aperture in the lower side of

20   the probe box. A solenoid coil with an oval cross-section is wound over the coil support. In the open-bore coil support, a support pad for a heel and external reference phantoms is inserted as will be depicted in **FIG.4** and **FIG.5**. An RF connector **14** is attached on the upper side of the probe box to supply RF currents for excitation

25   of proton spins and to receive NMR signals from proton spins.

An LC tank circuit is schematically depicted in **FIG.3**. In this figure, although the RF coil **15** is depicted as a solenoid with a circular

cross-section, the solenoid coil with the oval cross-section which is wound on the RF coil support **13** can be also used for the tank circuit. The capacitor **16** is used mainly for tuning of the tank circuit to the resonance frequency (8.9 MHz) and the capacitor **17** is used mainly for impedance matching to 50 ohms.

FIG.4 is a schematic cross-sectional depiction of a support pad **18** for a heel **19** and external reference materials denoted by **20** and **21**, inserted in the open-bore of the RF probe shown in **FIG.2**. For quantification of proton densities of bone marrow in calcaneus, to fix the heel and to acquire MRI images together with known reference materials such as water is essential, because tuning of the tank circuit varies person to person dependent on the size and position of the foot. In this figure, polystyrene foam is used for the support pad **18** and $CuSO_4$ water solution in cylindrical polyethylene bottles is used as the external reference material.

FIG.5 shows an alternative embodiment for the support pad **22**. A heel **23** is directly placed on the external reference material **24** which is flexible to fit any shape of a foot to improve the precision of RF field inhomogeneity correction.

FIG.6 is a standard 2D spin-echo pulse sequence used to quantify the proton density of bone marrow in calcaneus. Image intensity $I(x,y)$ acquired with the spin-echo sequence can be expressed except the J modulation effect of lipid as follows:

$$I(x,y) = kf(x,y)\rho(x,y)\{1 - p(x,y)\exp(-TR*/T_1(x,y))\}\exp(-TE/T_2(x,y)) , \quad (1)$$

where k is a constant, $f(x,y)$ represents spatial variation of image

intensity for a uniform sample, which is determined by inhomogeneity of static magnetic field, magnetic field gradients, RF magnetic field, and coil sensitivity (in principle the same as distribution of RF magnetic field), $p(x,y)$ represents a factor reflecting longitudinal

5   magnetization, which is unity for perfect 90 degree excitation over the cross-sectional plane, TR^* is a time close to the repetition time TR, $T_1(x,y)$ and $T_2(x,y)$ are T_1 and T_2 distributions in the plane.

FIG.7 shows a flow diagram of one method to quantify the proton density of bone marrow in calcaneus. To quantify the proton density

10   $\rho(x,y)$, spin-echo sequences with TR/TE=1200ms/12ms and TR/TE=1200ms/108ms are used for the heel measurements as denoted by **25**. Since the T_1 of the protons of the bone marrow in calaneus is about 250 ms at 8.9 MHz, MR image intensities at calcaneus acquired with the above sequences can be expressed as

15

$$I_1(x,y) = kf(x,y)\rho(x,y)\exp(-12/T_2(x,y)) \qquad (2)$$

$$I_2(x,y) = kf(x,y)\rho(x,y)\exp(-108/T_2(x,y)) . \qquad (3)$$

20   By using above two equations, $T_2(x,y)$ and $kf(x,y)\rho(x,y)$ are calculated as denoted by **26** and **27**. For the measurement of $kf(x,y)$, a $CuSO_4$ water solution phantom of which T_2 is about 43ms is placed in the RF coil instead of a heel and imaged with the same spin-echo sequences with TR/TE=500ms/12ms and TR/TE=500ms/48ms as denoted by **28**. By using

25   these images, $kf(x,y)$ is calculated when the proton density of water is defined as unity as denoted by **29** and **30**. The proton density $\rho(x,y)$ of the bone marrow when the proton density of water is defined as

unity is thus calculated by dividing $kf(x,y)\rho(x,y)$ by $kf(x,y)$ as denoted

by **31**. The trabecular bone volume fraction is then computed as

$1-\rho(x,y)$, as denoted by **32**.

FIG.8 and **FIG.9** schematically show cross-sectional images of a heel

5   and a water phantom with external reference phantoms. For actual

calculations for proton densities and T_2 relaxation times, mean

values in the squares depicted in **FIG.8** and **FIG.9** are used because

random noise is superimposed on the MR images and the precision is

limited.

10   **FIG.10** is a semilogarithmic plot of spin-echo image intensities of

a plant oil phantom measured against the echo time, showing the J

modulation effect on the spin-echo intensities. The similar property

is observed for the bone marrow in calcaneus because the bone marrow

has a similar chemical composition as the plant oil. The J modulation

15   effect, which is seen as the deviation of the spin-echo signal shown

from the straight line, is shown as an oscillation with about 100ms

period. Thus the T_2 decay of the lipid protons can be corrected

reasonably using the images acquired with TE=12ms and TE=108ms.

FIG.11 is a large flip-angle excitation 2D spin-echo pulse sequence

20   to shorten the repetition time for the quantification of the proton

density of bone marrow in calcaneus. By setting the excitation flip

angle to an angle between 90 and 180 degree, the recover of the

longitudinal magnetization becomes faster as depicted in **FIG.11**.

The arrows denoted by **33** to **41** show a time variation of longitudinal

25   nuclear magnetization.

While only a few specific exemplary embodiments of this invention

have been described in detail, those skilled in the art will readily

appreciate that many variations and modifications may be made in these exemplary embodiments while yet retaining many of the novel features and advantages of this invention. Accordingly, all such modifications and variations are intended to be included within the

5   scope of the appended claims.